



# Free-field Microphone Type MIC 01

## Calibration Chart

**Serial No:** AU0201086  
**Sensitivity:** -34.6 dB re 1V/Pa  
18.58 mV/Pa

**Valid at:**  
Temperature: 25 °C  
Ambient Static Pressure: 101.0 kPa  
Relative Umidity: 50 %  
Frequency: 1000 Hz  
Polarization Conditions: 8.2-5.6 V-KOhm

### Reference Instruments:

MANUF.	MODEL	SERIAL NO	CALIB. DATE
HP	3468A	2137A09681	20-06-2002
LARSON DAVIS	812	416	10-05-2002
B&K	4231	1759543	10-05-2002
B&K	4939	2345912	05-03-2002

**Date:** 06/07/02 **Certificate NO:** 9206 **Signature :** M.J.

